LR
ENVIRONMENTAL EQUIPMENT & TEST LAB

Humidity Test Profile For CSZ Chamber Model ZPHS-8-1.5-H/AC

EVENT FUNCTIONS/DIGITAL OUTPUT
EVENT 1 = CONDITIONING SYSTEM
EVENT 2 = HUMIDITY

User Profile Record

Copy this record and use it to plan profiles. Keep it with a Setup Page Parameter Record to document the controller's programmed settings.

Profile Name: LR HUMID

Date Programmed: _____

Programmed by: _____

Controller checked by: _____

X = ON
O = OFF

Step Nmbr	Step Type	Date/Day, Time	Wait for	Set Events 1	2	3	4	5	6	7	8	Time H M S	Rate	Set Pt 1	Set Pt 2	PID Set	Guar. Soak	Jump to Profile	Step	Repeats	End Step
1	RAMP		NO	X	X							5 S	-	25	20	1+6	NO	-	-	-	-
2	RAMP		NO	X	X							30 m	-	-	-	1+6	YES 1 NO 2	-	-	-	-
3	RAMP		NO	X	X							1 H	-	25	90	1+6	NO	-	-	-	-
4	SOAK		NO	X	X							2 H	-	-	-	1+6	YES 1 YES 2	-	-	-	-
5	RAMP		NO	X	X							1 H	-	60	95	1+6	NO	-	-	-	-
6	SOAK		NO	X	X							2 H	-	-	-	1+6	YES 1 YES 2	-	-	-	-
7	RAMP		NO	X	X							1 H	-	85	85	1+6	NO	-	-	-	-
8	SOAK		NO	X	X							2 H	-	-	-	1+6	YES 1 YES 2	-	-	-	-
9	RAMP		NO	X	X							1 H	-	40	90	1+6	NO	-	-	-	-
10	SOAK		NO	X	X							2 H	-	-	-	1+6	YES 1 YES 2	-	-	-	-
11	RAMP		NO	X	X							1 H	-	40	50	1+6	NO	-	-	-	-
12	SOAK		NO	X	X							2 H	-	-	-	1+6	YES 1 YES 2	-	-	-	-
13	RAMP		NO	X	X							1 H	-	25	20	1+6	NO	-	-	-	-
14	SOAK		NO	X	X							2 H	-	-	-	1+6	YES 1 NO 2	-	-	-	-
15	END		-			-						-	-	-	-	-	-	-	-	-	ALL OFF