LR
ENVIRONMENTAL EQUIPMENT & TEST LAB

For Quality Refurbished -Warranteed
Test Equipment Contact Us At:
323-770-0634 800-574-2748
sales@LRE.com www.LRE.com

Test Equity Fast Ramp
Temperature Chamber
Model #3007C

EVENT FUNCTIONS/DIGITAL OUTPUT
EVENT 1 (DIG OUT 1) CONDITIONER POWER
GUARANTEED SOAK (GSD) = 2.0

User Profile Record

Copy this record and use it to plan profiles. Keep it with a Setup Page Parameter Record to document the controller's programmed settings.

Profile Name: LR TEMP

Date Programmed: _____

Programmed by: 3007C

Controller checked by: _____

X = ON
O = OFF

Step Nmbr	Step Type	Date/Day, Time	Wait for	Set Events 1	2	3	4	5	6	7	8	Time H M S	Rate	Set Pt 1	Set Pt 2	PID Set	Guar. Soak	Jump to Profile	Step	Repeats	End Step
1	RAMP	–	NO	X								5 S	–	-70	–	1	NO	–	–	–	–
2	SOAK	–	NO	X								30 m	–	–	–	1	YES	–	–	–	–
3	RAMP	–	NO	X								5 S	–	150	–	1	NO	–	–	–	–
4	SOAK	–	NO	X								30 m	–	–	–	1	YES	–	–	–	–
5	RAMP	–	NO	X								5 S	–	-55	–	1	NO	–	–	–	–
6	SOAK	–	NO	X								2 H	–	–	–	1	YES	–	–	–	–
7	RAMP	–	NO	X								5 S	–	100	–	1	NO	–	–	–	–
8	SOAK	–	NO	X								2 H	–	–	–	1	YES	–	–	–	–
9	JUMP	–	–									–	–	–	–	–	–	LR TEMP	5	3	–
10	RAMP	–	NO	X								5 S	–	24	–	1	NO	–	–	–	–
11	SOAK	–	NO	X								24	–	–	–	1	YES	–	–	–	–
12	END	–	–									–	–	–	–	–	–	–	–	–	ALL OFF